

Kugelhahn G1.1/4" DN32 Zapfen/Muffe ZA 01/V AF1
Ball valve ZA 01/V AF1
 nr./no. 120403

Bajonettverschluss PS 11/N AF4
Bayonet coupling PS 11/N AF4
 nr./no. 561585

Zwischenrohr ZA 01/B AF1
Intermediate tube ZA 01/B AF1
 nr./no. 513596

Doppelnippel konisch 1.1/4"
Double nipple 1.1/4"
 nr./no.: 120865

Axialluftdüse PS 01/A AF2
Air purge PS 01/A AF2
 nr./no. 561553

Quarzscheibenvorsatz PS 01/I AF2
Quartz window PS 01/I AF2
 nr./no. 561487

Wärmefalle PS 01/K AF1
Thermal Insulating tube PS 01/K AF1
 nr./no. 513522

Pyrometer

Technische Daten

Axialluftdüse PS 01/A AF2
Air purge PS 01/A AF2

max .Druck 0,5 bar
pressure 0,5 bar max.

Es wird empfohlen ein Überdruckventil zu verwenden.
It's recommended to use a pressure relief valve.

Quarz-Scheibenvorsatz PS 01/I AF2
Quartz window PS 01/IA AF2

ca. 771 mm

Ø44,5 mm

M30x1,5

17,5 mm

15 mm

G1 1/4"

Schlauchtülle G1/8"

für Schlauch mit lichter Weite 8 bis 9 mm
Hose nozzle G1/8"
 for hose with inner width 8 to 9 mm

Seite 1 von 1
 Page 1 of 1

Gewicht: 3.210 Kg
 Weight: 3.210 Kg



- Headquarters
- Sales and Service Center
- Sales abroad



Keller HCW GmbH
Infrared Temperature Solutions (ITS)
Carl-Keller-Straße 2-10
49479 Ibbenbüren-Laggenbeck
Germany

www.keller.de/its
Tel. +49 (0) 5451 850
Fax +49 (0) 5451 85412
its@keller.de

Distributor



INOR Transmitter Oy
Unikkotie 13
FI-01300 Vantaa
Puhelin +358 10 4217900
Faksi +358 10 4217901
myynti@inor.fi